

महाराष्ट्र शासन
अधिष्ठाता यांचे कार्यालय, शासकीय आयुर्वेद रुग्णालय, वजिराबाद, नांदेड
OFFICE OF THE DEAN, GOVT. AYURVED HOSPITAL, VAZIRABAD, NANDED

Ph & Fax No. (02462) 232867

E-mail--gahnanded@gmail.com

जा.क्र. शाआरुनां/भांडार/ ४३२० /२०२१

दि. १०/१२/२०२१

प्रति,

विषय :- शासकीय आयुर्वेद रुग्णालय, नांदेड.

यंत्रसामुग्री/उपकरणे इ.पुरवठा करण्याकरीता दरपत्रक सादर करणे बाबत.

उपरोक्त विषयाच्या अनुषंगाने आपणांस कळविण्यांत येते की, जिल्हा वार्षिक योजना २०२१-२०२२ अंतर्गत मंजुर नियतव्ययातून शासकीय आयुर्वेद रुग्णालय, नांदेडचे बळकटीकरण तथा श्रेणीवर्धन याकरीता मंजुर निधीतून यंत्रसामुग्री/उपकरणे खरेदी करण्याकरीता दरपत्रके मागविण्यांत येत आहेत. सदर यंत्रसामुग्रीचे नांव, नग व विनिर्दिष्टे (स्पेशिफीकेशन) सोबत जोडण्यांत आले आहेत.

आपणांस सुचित करण्यांत येत की, रुग्णालयाकरीता आवश्यक यंत्रसामुग्री/उपकरणे इ.पुरवठा करण्याकरीता सोबत जोडण्यांत आलेल्या विनिर्दिष्टानुसार खालील नमुद यंत्रसामुग्रीचा पुरवठा करण्याकरीता कमीत कमी दराची मोहोरबंद पाकीटातील दरपत्रके दिनांक २४.१२.२०२१ रोजी सायं. ०५.०० वाजेपर्यंत पोहोचतील अशा बेताने अधिष्ठाता यांचे कार्यालयात खालील नमुद अटी व शर्ती च्या अधिन राहून सादर करण्यांत यावीत.

अ.क्र	बाब	कालावधी
1)	दरपत्रक सादर करण्याचा कालावधी	दिनांक १६.१२.२०२१ ते २४.१२.२०२१ (या कार्यालयीन कामकाजाच्या दिवशी व वेळेत, शासकीय व स्थानिक सुट्टीचे दिवस वगळून सकाळी १०.०० ते सायं. ०५.०० वाजेपर्यंत)
2)	दरपत्रक उघडण्याचा दिनांक व वेळ	दिनांक २७.१२.२०२१ रोजी सकाळी ११.०० वाजता.
3)	दरपत्रक उघडण्याचे ठिकाण	अधिष्ठाता शासकीय आयुर्वेद रुग्णालय, नांदेड यांचे दालन

अटी व शर्ती :-

- दिनांक २४.१२.२०२१ रोजी सायं. ०५.०० वाजल्यानंतर प्राप्त होणा-या दरपत्रकांचा विचार केला जाणार नाही.
- दरपत्रक सादर करतांना अशा प्रकारच्या खरेदीवर शासनाकडून आकारल्या जाणारे विविध कर (टॅक्सेस) स्वतंत्रपणे नमुद करावेत.
- दरपत्रक सादर करतांना जीएसटी क्रमांक नमुद असलेल्या लेटरहेडवर सादर करावेत.
- ज्या यंत्रसामुग्रीकरीता दरपत्रक सादर करण्यांत येत आहेत ते आय.एस.आय.चिन्हांकित असेल तर ठळकपणे नमुद करावे.
- दरपत्रक सादर करणा-या दरपत्रकाचा किंवा संस्थेचा दरपत्रक प्रक्रियेत भाग घेणा-या अन्य संस्थेमध्ये थेट स्वारस्य असू नये, तसे आढळल्यास दोन्ही किंवा अधिक दरपत्रके नाकारण्यात येतील. याकरीता सोबत जोडण्यांत आलेली हमीपत्र दरपत्रकधारकास सादर करणे बंधनकारक राहिल. (हमीपत्र विहित नमुन्यात सादर करावे.)

६. यंत्रसामुग्री खरेदी पश्चात यंत्रसामुग्री पुरवठा केल्यानंतर त्याचा देखभाल (वॉरंटी कालावधी) कालावधी ३ वर्षांचा असणे आवश्यक आहे असे हमी देण्यांत यावी.
७. दरपत्रका सोबत ज्या यंत्र सामुग्री करिता दरपत्रक सादर करण्यात येत आहे त्या यंत्रसामुग्रीची विस्तृत माहिती असलेली माहितीपत्रक (Product catalogue) जोडावे..
८. दरपत्रक सादर करतेवेळी मोहरबंद/सिलबंद पाकीटावर यंत्रसामुग्री करिताचे दरपत्रक असे स्पष्टपणे नमुद करावे.
९. खरेदी प्रक्रिया यशस्वीदरपत्रक स्विकारण्याच्यापुर्वी किंवा यशस्वी दरपत्रक स्विकारण्यात आल्यानंतर देखील कोणत्याहीवेळी लेखी नमुद करावयाच्या कारणामुळे रद्द करण्याचे अधिकार हे कार्यालय राखून ठेवीत आहे.
१०. विहित कालावधीत प्राप्त दरपत्रके दिनांक २७.१२.२०२१ रोजी सकाळी ११.०० वाजता अधिष्ठाता यांच्या दालनात उघडण्यात येतील.त्यावेळी कंपनीच्या एका प्रतिनिधीस स्वखर्चाने हजर राहण्याची मुभा राहिल.
- ११) दरपत्रक उघडल्यानंतर आपण सादर केलेल्या यंत्राच्या विनिर्दिष्टानुसार (स्पेशिफिकेशननुसार) सदर यंत्राची प्रत्यक्ष पडताळणी व आवश्यकता वाटल्यास चाचणी रुग्णालयातील संबंधीत तज्ञामार्फत करण्यांत येईल त्याकरीता आपण आवश्यक ती व्यवस्था करून ठेवावी.
- १२) पुरवठादाराने सदर खरेदी संदर्भातील आवश्यक ती माहिती या सोबतच्या ((Annexure-B) मध्ये सादर करावी.
- १३) प्रत्येक यंत्रसामुग्री/उपकरणे करिता स्वतंत्र दरपत्रके सादर करावीत.
- १४) खरेदीप्रक्रियेच्या अटी व शर्तीमध्ये कोणताही बदल करण्याचा अधिकार अधिष्ठाता शासकीय आयुर्वेद रुग्णालय,नांदेड हे राखून ठेवीत आहे.

(वैद्य वाय.आर.पाटील)

अधिष्ठाता

शासकीय आयुर्वेद रुग्णालय,नांदेड.

(2)

GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

Sr. No.	Name Instruments & Machinery	Specification	Qty.	Rate Per Unit
1	Reagent & Chemical Bottles of Borocillate	Made of borocillate glass size 100 to 200ml	1	
2	Glass Spirit Lamps for Practical Purpose	250ml spirit lamps with transparent glass make.	1	
3	Skin Ph Meter	<p>pH Range -2.00 to 16.00 pH</p> <p>pH Resolution 0.01 pH</p> <p>pH Accuracy ±0.02 pH</p> <p>pH Calibration automatic, at one or two points with two sets of buffers (standard 4.01, 7.01, 10.01 or NIST 4.01, 6.86, 9.18)</p> <p>Temperature Range -5.0 to 105.0°C / 23.0 to 221.0°F</p> <p>Temperature Resolution 0.1°C / 0.1°F</p> <p>Temperature Accuracy ±0.5°C (up to 60°C); ±1.0°C (outside) / ±1°F (up to 140°F); ±2.0°F (outside)</p> <p>Compensation automatic from -5.0 to 105.0°C (23 to 221°F)</p> <p>Electrode/Probe HI1414D/50 pre-amplified pH probe with internal temperature sensor, DIN connector and 1 m (3.3') cable (included)</p>	1	

(2)
GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

		<p>Battery Type/Life 1.5V (3) AAA / approximately 1200 hours of continuous use; auto-off after 8 minutes of non-use</p> <p>Environment 0 to 50°C (32 to 122°F); RH max 100%</p> <p>Dimensions 152 x 58 x 30 mm (6.0 x 2.3 x 1.2")</p> <p>Weight 205 g (7.2 oz.)</p> <p>Ordering Information HI99181 is supplied with HI1414D/50 flat tipped pH/temperature probe, HI70004 pH 4.01 buffer solution sachet, HI70007 pH 7.01 buffer solution sachet, HI700620 electrode cleaning and disinfection solution for skin residuals sachets (2), HI700621 electrode cleaning solution for skin grease and sebum sachets (2), batteries, with suitable Warranty upto 2 years</p>		
4	Fully automated computerized ECG machine (12 Channels)	<p>High-resolution 8" colour display</p> <p>Touch function keys</p> <p>Easy-to-clean sealed alphanumeric keyboard</p> <p>More than 8 hours battery capacity with ECG printout</p> <p>Resting rhythm up to 10 minutes</p> <p>Hook-up adviser with color-coded waveforms and anatomical model</p> <ul style="list-style-type: none"> • Lead reversal detection Fast and secure bidirectional Wi-Fi communication facility must be there. ETM including ETM Sport as well as CCAA Basic exercise software Work list Spirometry, based on the ultrasound technology of the Spiro Scout SP plus(Automatic and manual mode • Full size 8 1/2" x 11" • Integrated rechargeable battery lasts up to 3 hours of normal use • Single-button operation SCHILLER filter technology • Computerized ECG interpretation program including accurate measurements 	1	

(2)

GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

		<ul style="list-style-type: none">• RS-232 interface for data transmission to PC SEMA-200• Meets or exceeds the U.S. and international quality standards• Portable at only 4.8 kg / 11 lbs)		
5	BP Apparatus/Sphygmomanometer	Should be of mercurial type, having ISI mark with On/Off provision for reservoir, measuring range from 0 to 300mmHg, the mercury should be clean, double distilled & of 99.9% purity, the threading & fastening arrangement of the cuff should show no sign of slip or failure when subjected to the maximum test condition. It must provide cleaning brush to clean the manometer tube & a set of spare washers may be provided with each unit.	1	
6	Pediatric BP apparatus	Mercury BP apparatus with cuff size 4x8 cm, 6x12 cm, 9x18 cm	1	
7	Electric Plaster Cutter	220 volt/5 amp, AC Power, CPM 15000, Blade-Stainless steel, 125 watts, Dimensions-25.4x7.6x20.30CM	1	
8	Anesthesia Workstation	Physical- Screen Size-178mm (7") TFT, Bellows (Latex Free)-Adult bellow-20 to 1600ml Pediatric option-20 to 350ml Functional- Tidal Volume-20 to 1600ml, Breath Rate(BPM)-4 to 99 BPM IE ratio-1:0.5 to 1:9, Inspiratory Pressure-4 to 60cm H ₂ O Electronic PEEP-4 to 20cm H ₂ O Modes-Standby, Volume Controlled, Pressure Controlled, Spontaneous Spontaneous Mode-Active Volume and Pressure Alarms Fresh Gas Compensation-Automatic Tidal Volume Adjustment Inspiratory Pause-25% of Insp. Time., Power-200-240 VAC, 50-60 Hz, 0.3Amax Battery Backup-2+hours Drive Gas-Oxygen or Air (Dry, Oil free) at 45 to 100 PSI(310 to 690 Kpa) Oxygen Monitor-Fuel Cell type (21% & 100%) Ventilator Performance accuracy of delivered volumes < 100ml + or - 50%. 100 to 300ml + or - 20% > 300ml + or - 10% Display Parameters-Delivered Tidal Volume, Airway Pressure, FiO ₂ %, Battery Status, Alarms. Alarm-User Adjustable-Low & High Tidal Volume Alarms-High & Low TV, Low & High O ₂ Concentration, Patient Disconnection, High AW	1	

(2)
GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

		Pressure, Mains failure, Low battery & Low Drive gas Low and High O ₂ concentration. Optional -Digital Flow Meter Assembly. Vaporizer- Sevoflurane		
9	Instrument Cabinet O.T.	Width-610mm Depth-355mm Height-1525mm Frame-CRCA rectangular tube & CRCA sheet panels Five Shelves, door & two sides of 5mm thick fine glass Pre-treated and epoxy powder coated.	1	
10	Double Foot Steps	Width-355 mm, Height-457mm, Length-400mm 25mm CRCA Tubes with Aluminium Chequered Strips Pre- treated & Epoxy Powder Coated	1	
11	पंचकर्म मॉडेल	1.Kuti swed , 2. jatanak swed , 3.kuti praveshika 4.dhupan yantra, 5.Kumbhi swed, 6.prastar swed, 7. holak swed, 8.kup swed , 9.karshu swed, 10. parishek swed, 11. ashmaghan swed, 12. murdha tailam-pichu, 13.shirobasti, 14.shirodhara, 15.(parishek) , 16. swedan peti, (sleeping position), 17. swedan peti, (sitting position).	1	

(2)

GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

	<p>18.droni(dharapatri), 19.karana puran, 20.vrana basti, 21. vrana dhupan, Raktamokshan : 22.prachhan, 23.shiravedh, 24.jalouka, 25. shrung, 26.alabu, 27.ghati, 28 Ideal position of basti , 29. Ideal position of uttar basti (male) 30. Ideal position of uttar basti(female), 31Bhu swed, 32.Trarpan, 33. swed yantra,</p>	1	
12	Basic Steam 3 KW	a) Stainless steel boiler duly insulated and housed in stainless steel cabinet with buff finish. b) Long life, low watt density, copper immersion heater with faster and efficient heating system. c) Electronically controlled automatic water feed with 'Low' or 'No' water cut - off system. d) Double thermostatic control with pressure safety arrangement. e) Continuous steam generation, even during water feeding and temperature cut-off time. f) Remote control panel with silver tipped contractors for millions of trouble free 'ON-OFF' operations. g) 1/2" dia solinade valve for cut off steam in cylindrical operational unit. h) Chrome plated brass fittings with special in line filter with double non return valve system. Imported thermostat timer and indicating lamps	1
13	Basic Steam 4 KW	i) Stainless steel boiler duly insulated and housed in stainless steel cabinet with buff finish. j) Long life, low watt density, copper immersion heater with faster and efficient heating system. k) Electronically controlled automatic water feed with 'Low' or 'No' water cut - off system. l) Double thermostatic control with pressure safety arrangement.	1

(2)
 GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
 PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

		<p>m) Continuous steam generation, even during water feeding and temperature cut-off time. n) Remote control panel with silver tipped contractors for millions of trouble free 'ON-OFF' operations. o) 1/2" dia solinade valve for cut off steam in cylindrical operational unit. p) Chrome plated brass fittings with special in line filter with double non return valve system. Imported thermostat timer and indicating lamps</p>		
14	Mooshs (Crucibles) Diff. Size)	Medium (4 No.)		1
15	Pressure Cooker	10 Liter		1
16	Charakokt Swedan prakar Model	Portable Acrylic models 15"x15"x15"		1
17	Trigarbha Rasayan Kuti	Portable Acrylic models 15"x15"x15"		1
18	Foetal doppler	dimensions 216*90*205mm (L*B*H), wt approx 1.9 kg, rechargeable and dry battery operating system		1
19	BMI Calculator Machine	Omron HBF 375 Karada Scan Complete Digital Body Composition Monitor With 3 Months Memory To Monitor BMI, Segmental Body Fat & Skeletal Muscle,		1
20	Sound system	BOSE SoundLink revolve 2 portable bluetooth speaker wireless water resistant speaker		1
21	Camcorder camera With Tripod stand	SONY HDR-CX 405 CAM CORDER CAMERA BLACK		1
22	Collar mike	SONY VCT-R640 TRIPOD BLACK		
23	Models of Naturopathy equipment's	JBL Commercial CSLM20B omnidirectional lavalier Set of 12 models made up of plastic Manufactured by Jemini Fiber glass		1
				1

(2)
 GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
 PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

<p>24</p> <p>Trinocular Research Microscope</p>	<p>OPTICAL SYSTEM Infinity corrected plan optics with Uniformly centered, interchangeable & parfocal , Tropicalized anti fungus treatment ensures image excellence for long periods in conditions favouring to fungus growth.</p> <p>OBJECTIVES Plan Infinite Achromat 4X, 10X, 40X, 100X</p> <p>EYEPIECES WH 10x (FN 20mm) paired eyepiece with eyeguards . The unique optical design provides relief from eye fatigue and renders wide-field images of utmost clarity. Compatible with optionally available eyepiece micrometer.</p> <p>DIOPTR CORRECTION Dioptic adjustment available on the left eyepiece tube</p> <p>OBSERVATION HEAD Binocular head (30 degree inclined seidentopf), 360 degree rotatable, with diopter adjustment on the left . Interpupillary distance 48-75mm</p> <p>Trinocular head (30 degree inclined seidentopf), 360 degree rotatable, with diopter adjustment on the left . Interpupillary distance 48-75mm.</p> <p>POWER SUPPLY (a) Built-in illumination base with pre-centered 6V 20W halogen light source coupled with an efficient collector lens system. Universal power supply 100V-240V AC 50/60Hz single phase. (b) 3W LED light source high brightness, longlife (30,000hrs) (c) LED light source (with battery back-up) high brightness, longlife (30,000hrs). Battery back-up in-built NiMH rechargeable batteries.</p> <p>NOSE PIECE</p>	<p>1</p>
---	---	----------

(2)
 GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
 PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

	<p>Quadruple revolving outward nosepiece based on precision ball-bearing mechanism with positive click stop.</p> <p>BODY Aluminium die- cast body with all critical movements based on ball bearing and wire guides thereby ensuring smooth and precise manipulation</p> <p>MECHANICAL STAGE Co-axial low drive mechanical rackless stage 120 mm x 132 mm with traverse area of 76 mm x 30 mm with single slide holder.</p> <p>FOCUSING SYSTEM Co-axial coarse and fine controls with a focus adjustment and fine adjustment knobs. Coarse focus range 20mm. Fine focus rotation 0.25mm</p> <p>CONDENSER HOLDER Rack and pinion mounted condenser holder</p> <p>CONDENSER Pre-centered abbe condenser with aperture iris diaphragm (N.A. 1.25) focusable with rack & opinion and a continuously variable iris diaphragm with a removable blue filter for day light observation</p> <p>MAGNIFICATION RANGE 40x ~ 1000x for observation</p>	
25	Digital Camera for Research Microscope with CMOS sensors	1 With an Aluminium alloy housing and standard C Mount with USB 2.0 interface, ideal pathology & research applications 14 Megapixels and fully compatible with Windows CAMERA CASING-Metal Alloy, IMAGE SENSOR-CMOS 14 megapixel, PIXEL SIZE-1.4 x 1.4 microns, RESOLUTION (MAXIMUM)-4096 x 3288 pixels, EXPOSURE TIME-0.4 ms to 2000 ms, SPECTRA

(2)
GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

26	Low profile Disposable blades for microtome	RANGE-380-650 nm with IR Filter. Type-Low Profile Material-Stainless Steel Usage/Application-Histology Type of Cutting Edge-Angled Quantity Per Pack-50 Piece	1	
27	High profile Disposable blades for microtome	Type-High Profile Material-Stainless Steel Usage/Application-Histology Type of Cutting Edge-Angled Quantity Per Pack-50 Piece	1	
28	Pre coated aluminium Plates for Chromatography	TLC Plates Aluminium Backed Silica Gel 60 F254 2.5 X 7.5cm (80/Box)	1	
29	Digital-Microprocessor pH Meter	During the pH mode. Up to 5 points push-button calibration with auto- buffer recognition. Selectable pH buffer standards and temperature units. Automatically calculates and displays electrode slope after calibration. Automatic Temp. Compensation (ATC) Calibration due reminder prompts Absolute or Relative millivolt (R.mV) measurement . Automatic endpoint function Setup menu should let user customizes buffer standards, calibration points, stability conditions, resolution, temperature units, date and time to meet measurement requirements. Expanded memory up to 500 readings. Built-in real-time clock to meets GLP standards. Stored data should be transferred into computer via USB communication interface.	1	
30	Electronic Analytical Balance	there should be three support points, two of which are adjustable to provide leveling. The base contains a high grade aluminum sensor, which gives clear and concise weighing results. protected by a three door glass draft shield surrounding the 43.2" diameter platter. The milligram (0.001 g) range should be fast, accurate and suitable for many different applications in the laboratory, including counting and percentage weighing. Communication via a serial RS232 port to provide GLP compliance and data collection. The capacity of the balance offers accurate weighing up to 300 g x 0.001 g, and reads in 5 weighing units.		
31	Micropipette	2-20µl C3 Single Channel Fully Autoclavable Variable Volume Pipette,	1	
32	Rotary Shaker	Digital Speed Indicator.	1	

(2)

GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

		<p>Digital Countdown Timer: 99 min. Shaking Orbit: 10 mm. Flask Capacity: 5x500 m Plate Size: 12x12 in Phase: Single Warranty :12 Months against any manufacturing defects Voltage 220-240 V Frequency 50 Hz Dimension 320x415x130 mm Speed Range 50-400 rpm</p>		
33	UV chamber TLC Visualization	<p>Size / dimension: 15 X 12 X10 Light source: UV lamp Material: SS Automation Grade: Manual Power Source :230</p>	1	
34	Laboratory hot air Oven	<p>Double Wall fabrication Two/ Three removable shelves Elegant Door closing Nii-chrome Wire Heating element L-shaped Thermometer Control Panel Inner Chamber Size: 305 X 305 X 305 mm Capacity (Litre): 20 Voltage: 220 V Type of Product: Laboratory Oven</p>	1	
35	Glass Jars with Steel Lid	Transparent glass jars of capacity 5 litres with broad opening and air tight steel lid in cylindrical shape	1	
36	TLC chamber	Twin trough chamber for TLC plates developing of size 20 X 20	1	
37	Steel Racks with height adjustable compartments	<p>Steel racks 3'x6'x1' with height adjustable compartments Five Compartments with 20 Gauge/ CR Sheet Adjustable height ,Backside Covered with Sheet/ Steel Gauge 22, adjustable Frame 1 1/2" – 3mm</p>	1	

(2)

GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022


38	College Name Board(120'x5')	- 4/100 feet Board Size - Aluminium coil /LED Letter size - 3.5 Feet Hight /100 feet Length 3. 5 X 100 = 350 (Ads company LED 3 Year Warranty) - Flexi Bond Panel 4" Feet Height / 100 Feet Length 4 "X 100 = 400 (10 Year Garrenty) MS Work Power supply & Garenty 1 Year - Board Maintance 3 Year free.	1	
39	Chest Stand	Moveble(15-20kg) for X Ray film shooting	1	
40	Revolving stools	.Top width-280mm, Depth-280mm Height-445mm Frame-25mmX 25 mm CRCA square tube Pre-treated & Epoxy Coated Stainless steel Top	1	
41	Chairs for Doctors	Length(40 inches),Width (20 inches),Height (27 inches)Primary material-Fabric Colour- black, Style-Modern,Cushionted Seat & Back, Torsion Knob-Tilt Lock,Padded Arm Rest,Durable & Long Lasting	1	
42	Analog type Stop Watches	Preferably Diamond Make – 1/5 sec Resolution/Least Count: 1/5 sec – 0.2 sec Dial: 60 Seconds Max Time: 30 Minutes Made of Stainless Steel Single button construction. Single button for Start->Stop->Reset	1	
43	Emergency Crash Cart O.T.	Length-840mm Depth-400mm Height-1640mm Stainless Steel Square Frame Top tray for monitor & pulse oximeter, six removable bins Two modular storage units of three drawers each one with lock Castors 125mm,2 with brake.	1	

(2)

GOVT AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS AND MACHINERY UNDER (D.P.D.C) BUDGET 2021-2022

Oxygen cylinder Holder, I.V Rod, Cardiac Massage Board.

44	कुकर		5 lit. steel cooker which will run on induction too.	1
45	अग्नीकर्म शलाका		पंचधातू युक्त	1
46	कपींग सेट		Hizama cupping sets	1
47	Chairs for Teaching Staff		Length(40 inches),Width (20 inches),Height (27 inches)Primary material-Fabric Colour- black,Style-Modern Cushioned Seat & Back,Torsion Knob-Tilt Lock,Padded Arm Rest,Durable & Long Lasting	1
48	Sliding Record Keeping Book Case		Size - 78" X 36" X 18" Five Compartment with fixed shelved CR Sheet, Steel Gauge 20 Powder coated paint premium Sliding movable 4 bearing on Track at base Inter Locking system to join 4 cupboards. Locking System Handle on sides. Transportation and installation Cost should include all Taxes local/GST etc Product Warranty of three Years.	1
49	Sliding Record Keeping Cabinets Case		Size - 78" X 36" X 18" Five Compartment with fixed shelved CR Sheet, Steel Gauge 20 Powder coated paint premium Sliding movable 4 bearing on Track at base Inter Locking system to join 4 cupboards. Locking System Handle on sides Transportation and installation Cost should include all Taxes local/GST etc Product Warranty of three Years.	1


DEAN
GOVT. AYURVED HOSPITAL
NANDED.

सहपत्र (Annrxure)- A
शासकीय आयुर्वेद रुग्णालय, नदिड.

SPECIFICATION OF.....

.क्र. दरपत्रकाव्दारे आवश्यक असलेले विनिर्दिष्टे Quotation Specification as asked in the quotation	पुरवठादाराने प्रस्तावित केलेली तत्सम विनिर्दिष्टे कंपनीच्या नावासहित Equivalent Specifications quoted by the supplier with the name of the manufacturer	पुरवठादाराने नमूद केलेली विनिर्दिष्टे दरपत्रकान्वये विचारणा करण्यांत आलेल्या विनिर्दिष्टानुसारच आहेत किंवा नाही. प्रत्येक बाबीसमोर होय किंवा नाही स्पष्टपणे नमूद करावे. Whether the Specification quoted by the supplier is as per specification asked for indicated against each item clearly YES/NO	पुरवठादाराने नमूद केलेली विनिर्दिष्टे दरपत्रकान्वये विचारणा करण्यांत आलेल्या असल्यास प्रत्येक बाबीसमोर स्पष्टपणे नमूद करावे. If the Specification quoted is not as per specification then variation /deviation should be clearly indicated against each item asked for.
१.			

Annrxure-B

कंपनीच्या लेटरहेडवर खालील नमूद तपशिलाप्रमाणे माहिती भरणे .

प्रति,
अधिष्ठाता,
शासकीय आयुर्वेद रुग्णालय,
वजिराबाद,
नांदेड.

वार्षिक योजना २०२१-२०२२

.....या यंत्रसामुग्री करिताचे दरपत्रक

अ.क्र.	तपशिल	आवश्यक माहिती वा प्रमाणपत्रे जोडण्यात आलेली आहेत किंवा कसे
१.	पुरवठादाराचे नांव	
२.	कंपनीचा पत्ता,दूरध्वनी ई-मेल आयडीसहित	
३.	कंपनीचा नोंदणी क्रमांक (Registration No.)	
४.	यंत्राचे नांव आणि मॉडेल क्रमांक	
५.	विनिर्दिष्टे (सोबत जोडण्यात आलेल्या सहपत्रानुसार) Annexure-A	
६.	GST No.	
७.	शासनाचे कर	
८.	इतर कर	
९.	I.S.I. Mark (आयएसआय चिन्हांकित आहे किंवा कसे असल्यास तत्संबंधीचे प्राधिकरणाकडील प्रमाणपत्रांची छायांकित प्रत जोडण्यात आली आहे किंवा कसे)	
१०.	उत्पादक असल्यास तसे नमूद करावे अथवा प्राधिकृत पुरवठादार असल्याबाबतचे उत्पादक कंपनीचे पत्र	
११.	सदर यंत्राचा इतर शासकीय /निमशासकीय कार्यालयास पुरवठा केला असल्यास अशा कार्यालयाचे पत्ते व दूरध्वनी क्रं.अथवा पुरवठा आदेशांची छायांकित प्रत.	
१२.	हमी कालावधी (Warranty Period)	
१३.	यंत्राची माहिती पुस्तिका/पत्रक	
१४.	हमीपत्र	
१५.	इतर अटी व शर्ती असल्यास	

स्थळ :-

दिनांक :-

पुरवठादाराचा अधिकृत स्वाक्षरी व शिक्कांसहित

हमीपत्र -१ (Undertaking)

मी M/s

पत्ता :-

लेखी हमी देत आहे की, आपल्या कार्यालयामार्फत

..... खरेदी करणेसाठी राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी भाग घेत असून माझे किंवा माझ्या संस्थेचे खरेदी प्राधिकारी / खरेदी समिती वा कार्यालयाशी कोणत्याही प्रकारचे हितसंबंध / संघर्ष नाहीत तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ :-

पुरवठादारांच्या अधिकृत स्वाक्षरी व शिक्कासह

दिनांक :-

हमीपत्र -२ (Undertaking)

मी M/s

पत्ता :- लेखी हमी देत आहे की,

आपल्या कार्यालयामार्फत खरेदी करणेसाठी राबविण्यात

येत असलेल्या दरपत्रक प्रक्रियेत मी/माझी संस्था केवळ एकच दरपत्रक सादर करित आहे. त्याचप्रमाणे दरपत्र

प्रक्रियेत भाग घेणाऱ्या अन्य संस्थासोबत माझा/माझ्या संस्थेचा कोणत्याही प्रकारचा प्रत्यक्ष /अप्रत्यक्ष संबंध वा सहभाग

नाही तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ :-

पुरवठादारांच्या अधिकृत स्वाक्षरी व शिक्कासह

दिनांक :-