

महाराष्ट्र शासन
अधिष्ठाता यांचे कार्यालय, शासकीय आयुर्वेद महाविद्यालय, वजिराबाद, नांदेड
OFFICE OF THE DEAN, GOVT. AYURVED COLLEGE, VAZIRABAD, NANDED

Ph & Fax No. (02462)234026

E-mail--gac.nanded@gmail.com

जा.क्र. शाआमनां/लेखा/डे-२/ 243 /२०२१

दि. 14 / 1 / २०२१

प्रति,

विषय :- शासकीय आयुर्वेद महाविद्यालय, नांदेड.

यंत्रसामुग्री/उपकरणे इ.पुरवठा करण्याकरीता दरपत्रक सादर करणे बाबत.

उपरोक्त विषयाच्या अनुषंगाने आपणांस कळविण्यांत येते की, वार्षिक योजना २०२०-२०२१ अंतर्गत मंजूर नियतव्ययातून शासकीय आयुर्वेद महाविद्यालय तथा संलग्नित रुग्णालय, नांदेडचे बळकटीकरण तथा श्रेणीवर्धन याकरीता मंजूर निधीतून यंत्रसामुग्री/उपकरणे खरेदी करण्याकरीता दरपत्रके मागविण्यांत येत आहेत. सदर यंत्रसामुग्रीचे नांव, नग व विनिर्दिष्टे (स्पेशिफीकेशन) सोबत जोडण्यांत आले आहेत.

आपणांस सुचित करण्यांत येते की, महाविद्यालयाकरीता व रुग्णालयाकरीता आवश्यक यंत्रसामुग्री/उपकरणे इ.पुरवठा करण्याकरीता सोबत जोडण्यांत आलेल्या विनिर्दिष्टानुसार खालील नमुद यंत्रसामुग्रीचा पुरवठा करण्याकरीता कमीत कमी दराची मोहोरबंद पाकीटातील दरपत्रके दिनांक २८.०१.२०२१ रोजी सायं. ०५.०० वाजेपर्यंत पोहोचतील अशा बेताने अधिष्ठाता यांचे कार्यालयात खालील नमुद अटी व शर्ती च्या अधिन राहून सादर करण्यांत यावीत.

अ.क्र	बाब	कालावधी
1)	दरपत्रक सादर करण्याचा कालावधी	दिनांक १५.१.२०२१ ते २८.०१.२०२१ (या कार्यालयीन कामकाजाच्या दिवशी व वेळेत शासकीय व स्थानिक सुट्टिचे दिवस वगळून सकाळी १०.०० ते सायं. ०५.०० वाजेपर्यंत)
2)	दरपत्रक उघडण्याचा दिनांक व वेळ	दिनांक २९.०१.२०२१ रोजी सकाळी ११.०० वाजता.
3)	दरपत्रक उघडण्याचे ठिकाण	अधिष्ठाता शासकीय आयुर्वेद महाविद्यालय, नांदेड यांचे दालन

अटी व शर्ती :-

- दिनांक २८.०१.२०२१ रोजी सायं. ०५.०० वाजल्यानंतर प्राप्त होणा-या दरपत्रकांचा विचार केला जाणार नाही.
- दरपत्रक सादर करतांना अशा प्रकारच्या खरेदीवर शासनाकडून आकारल्या जाणारे विविध कर (टॅक्सेस) स्वतंत्रपणे नमुद करावेत.
- दरपत्रक सादर करतांना जीएसटी क्रमांक नमुद असलेल्या लेटरहेडवर सादर करावेत.
- ज्या यंत्रसामुग्रीकरीता दरपत्रक सादर करण्यांत येत आहेत ते आय.एस.आय.चिन्हांकित असेल तर ठळकपणे नमुद करावे.
- दरपत्रक सादर करणा-या दरपत्रकाचा किंवा संस्थेचा दरपत्रक प्रक्रियेत भाग घेणा-या अन्य संस्थेमध्ये थेट स्वारस्य असू नये, तसे आढळल्यास दोन्ही किंवा अधिक दरपत्रके नाकारण्यात येतील. याकरीता सोबत जोडण्यांत आलेली हमीपत्र दरपत्रकधारकास सादर करणे बंधनकारक राहिल. (हमीपत्र विहित नमुद न्यात सादर करावे.)

GACN/2018-2019/D-2

६. यंत्रसामुग्री खरेदी पश्चात यंत्रसामुग्री करीता देण्यात येणारा हमी कालावधी (वांरटी कालावधी) ठळकपणे नमुद करावा
७. दरपत्रका सोबत ज्या यंत्र सामुग्री करिता दरपत्रक सादर करण्यात येत आहे त्या यंत्रसामुग्रीची विस्तृत माहिती असलेली माहितीपत्रक (Product catalogue) जोडावे..
८. दरपत्रक सादर करतेवेळी मोहरबंद/सिलबंद पाकीटावर यंत्रसामुग्री करिताचे दरपत्रक असे स्पष्टपणे नमुद करावे.
९. खरेदी प्रक्रिया यशस्वीदरपत्रक स्विकारण्याच्यापुर्वी किंवा यशस्वी दरपत्रक स्विकारण्यात आल्यानंतर देखील कोणत्याहीवेळी लेखी नमुद करावयाच्या कारणामुळे रद्द करण्याचे अधिकार हे कार्यालय राखून ठेवीत आहे.
१०. विहित कालावधीत प्राप्त दरपत्रके दिनांक २९.०१.२०२१ रोजी सकाळी ११.०० वाजता अधिष्ठाता याच्या दालनात उघडण्यात येतील.त्यावेळी कंपनीच्या एका प्रतिनिधीस स्वखर्चाने हजर राहण्याची मुभा राहिल.
- ११) दरपत्रक उघडल्यानंतर आपण सादर केलेल्या यंत्राच्या विनिर्दिष्टानुसार (स्पेशिफिकेशननुसार) सदर यंत्राची प्रत्यक्ष पडताळणी व आवश्यकता वाटल्यास चाचणी महाविद्यालयातील संबंधीत तज्ञामार्फत करण्यांत येईल त्याकरीता आपण आवश्यक ती व्यवस्था करून ठेवावी.
- १२) पुरवठादाराने सदर खरेदी संदर्भातील आवश्यक ती माहिती या सोबतच्या ((Annexure-B) मध्ये सादर करावी.
- १३) प्रत्येक यंत्रसामुग्री/उपकरणे करीता स्वतंत्र दरपत्रके सादर करावीत.
- १४) खरेदीप्रक्रियेच्या अटी व शर्तांमध्ये कोणताही बदल करण्याचा अधिकार अधिष्ठाता शासकीय आयुर्वेद महाविद्यालय,नांदेड हे राखून ठेवीत आहे.



(वैद्य वाय.आर.पाटील)

अधिष्ठाता

शासकीय आयुर्वेद महाविद्यालय,नांदेड.

सहपत्र (Annrxure)- A
शासकीय आयुर्वेद महाविद्यालय, नांदेड.

SPECIFICATION OF.....

अ.क्र.	दरपत्रकाव्दारे आवश्यक असलेले विनिर्दिष्टे Quotation Specification as asked in the quotation	पुरवठादाराने प्रस्तावित केलेली तत्सम विनिर्दिष्टे कंपनीच्या नावासहित Equivalent Specifications quoted by the supplier with the name of the manufacturer	पुरवठादाराने नमूद केलेली विनिर्दिष्टे दरपत्रकान्वये विचारणा करण्यांत आलेल्या विनिर्दिष्ट्यानुसारच आहेत किंवा नाही. प्रत्येक बाबीसमोर होय किंवा नाही स्पष्टपणे नमूद करावे. Whether the Specification quoted by the supplier is as per specification asked for indicated against each item clearly YES/NO	पुरवठादाराने नमूद केलेली विनिर्दिष्टे दरपत्रकान्वये विचारणा करण्यांत आलेल्या विनिर्दिष्टे यामध्ये काही बदल वा तफावत असल्यास प्रत्येक बाबीसमोर स्पष्टपणे नमूद करावे. If the Specification quoted is not as per specification then variation /deviation should be clearly indicated against each item asked for.
१.				

Annrxure-B

कंपनीच्या लेटरहेडवर खालील नमूद तपशिलाप्रमाणे माहिती भरणे.

प्रति,
अधिष्ठाता,
शासकीय आयुर्वेद महाविद्यालय,
वजिराबाद,
नांदेड.

वार्षिक योजना २०२०-२०२१

.....या यंत्रसामुग्री करिताचे दरपत्रक

अ.क्र.	तपशिल	आवश्यक माहिती वा प्रमाणपत्रे जोडण्यात आलेली आहेत किंवा कसे
१.	पुरवठादाराचे नांव	
२.	कंपनीचा पत्ता,दूरध्वनी ई-मेल आयडीसहित	
३.	कंपनीचा नोंदणी क्रमांक (Registration No.)	
४.	यंत्राचे नांव आणि मॉडेल क्रमांक	
५.	विनिर्दिष्टे (सोबत जोडण्यात आलेल्या सहपत्रानुसार) Annexure-A	
६.	GST No.	
७.	शासनाचे कर	
८.	इतर कर	
९.	I.S.I. Mark (आयएसआय चिन्हांकित आहे किंवा कसे असल्यास तत्संबंधीचे प्राधिकरणाकडील प्रमाणपत्रांची छायांकित प्रत जोडण्यात आली आहे किंवा कसे)	
१०.	उत्पादक असल्यास तसे नमूद करावे अथवा प्राधिकृत पुरवठादार असल्याबाबतचे उत्पादक कंपनीचे पत्र	
११.	सदर यंत्राचा इतर शासकीय /निमशासकीय कार्यालयास पुरवठा केला असल्यास अशा कार्यालयाचे पत्ते व दूरध्वनी क्रं.अथवा पुरवठा आदेशांची छायांकित प्रत.	
१२.	हमी कालावधी (Warranty Period)	
१३.	यंत्राची माहिती पुस्तिका/पत्रक	
१४.	हमीपत्र	
१५.	इतर अटी व शर्ती असल्यास	

स्थळ :-

दिनांक :-

पुरवठादाराचा अधिकृत स्वाक्षरी व शिक्कांसहित

GACN/2018-2019/D-2

हमीपत्र -१ (Undertaking)

मी M/s.

पत्ता :-
लेखी हमी देत आहे की, आपल्या कार्यालयामार्फतखरेदी
करणेसाठी राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी भाग घेत असून माझे किंवा माझ्या
संस्थेचे खरेदी प्राधिकारी/खरेदी समिती वा कार्यालयाशी कोणत्याही प्रकारचे हितसंबंध/संघर्ष
नाहीत तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ :-

पुरवठादारांच्या अधिकृत स्वाक्षरी व शिक्कासह

दिनांक :-

हमीपत्र-२ (Undertaking)

मी M/s.

पत्ता :- लेखी
हमी देत आहे की, आपल्या कार्यालयामार्फत
खरेदी करणेसाठी राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी/माझी संस्था केवळ एकच
दरपत्रक सादर करित आहे. त्याचप्रमाणे दरपत्रक प्रक्रियेत भाग घेण्या-या अन्य संस्थांसोबत
माझा/माझ्या संस्थेचा कोणत्याही प्रकारचा प्रत्यक्ष/अप्रत्यक्ष संबंध वा सहभाग नाही तसे आढळून
आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ :-

पुरवठादारांच्या अधिकृत स्वाक्षरी व शिक्कासह

दिनांक :-

शासकीय आयुर्वेद महाविद्यालय, नांदेड

वार्षिक योजना सन 2020-2021 महाविद्यालय व रुग्णालय बळकटीकरण व श्रेणीवर्धन अंतर्गत यंत्रसामुग्री/उपकरणे खरेदी ३ लाखा पेक्षा कमीचा प्रस्ताव

अ.क्र.	यंत्रसामुग्रीचे नांव	विनिर्दिष्टे	मागणी नग संख्या
1	Western Green's Pipette For ESR	Blood Sedimentation Tubes of Borocillate glass Suitable for determining sedimentation rates ,the values stated in SI units.	40
2	Haematocrit Tube	Blood Sedimentation Tubes of Borocillate glass Suitable for determining sedimentation rates & Volume of Packed cells ,the values stated in SI units. regarded as standards.	40
3	Stop watches	Preferable any standard make for laboratory use preferred Diamond make	94
4	Balances	Preferable any standard balance for laboratory use. 1) Wooden Analytical Balance Scale, 2) Capacity--220gm, 3) Pan Size-3.5", 4) Readability--0.1gm, 5) Automation Grade-Manual, 6) Frame Material--Wooden, 7) Width-18 Inches, 8) Depth-11Inches, 9) Height--17 Inches	1
5	Stethoscopes	measuring tip: approx. 16.5 cm earpieces made of soft plastic total length: approx. 1 m tight ear clips - ear clips adapted to the human head- for preventive maintenance and servicing	158
6	Clinical Thermometer	Preferable any standard make for laboratory use.	100
7	Knee Hammer	Preferable any standard make for laboratory use.	102
8	Tuning Forks	Preferable any standard make for laboratory use.	104
9	Sterilizer	Preferable any standardmake for laboratory use.	3
10	Pyrometer	Range 1200 °C Infrared	02
11	Thermo meter	Mercury; temp range 0 to 360°C	02
12	Digital PH meter	pH Measuring Range – 0.00 – 14.00, Resolution – 0.01 pH, Accuracy - ±0.03 pH	02
13	Tab Friability apparatus	2 drums, 20-60 RPM- Drum Rotation, 220 V AC 50 Hz	01
14	Projector	SVGA projector Brightness; 3300lm with HDMI Port	01

15	Digital Viscometer	Digital viscometer	01
16	Water Bath	Water bath with Thermo-regulator, Stainless steel body, Bath capacity – 15L, Digital Display	02
17	Tablet Hardness Tester	Monsanto, Stailless steel body, scale – 0-10 kg,	01
18	Bubble CPAP	Continues positive airway pressure therapy Unit with fully automatic Servo humidifier, Medical grade low noise air compressor with high flow safety protection Inbuilt low flow Air/Oxygen blender to set FiO2 percentage (21% -100% with flow meter to set output flow) , Reusable unique design bubble generator with pressure gauge to set airway pressure, Infant heated wired breathing circuit, IV stand, utility basket, anti skid lockable casters. universal infant nasal prong, septal width Nose, Diameter nasal prong (3 mm/2mm), (3.5mm/2mm), (4mm/3mm), (4.5mm/4mm), (5mm/4mm), (5mm/5mm), (5.5mm/6mm), (6.5mm/7mm)-10 each Infant nasal mask-small, medium, large Infant nasal tubing-Standard size-50 mm, 70 mm, 100mm Universal size-50 mm, 70 mm, 100mm	01
19	Pulse Oxymeter Fingure	ECG, NIBP, Temperature, spo2, Pulse Rate	06
20	Pulse Oxymeter multipara	ECG, NIBP, Temperature, spo2, Pulse Rate	01
21	Nebulizer	Paediatric and Adult with Face mask	15
22	Syringe Pump	Syringe size-2ml, 5ml, 10ml, 20ml, 50ml	02
23	Electronic Pediatric Weighing Machine	To measure weight up to 100 kg	01
24	Stadiometer Pediatric	To measure height from '0' cm to 180 cm	02
25	Oxygen Flow meter	Size : 61 x 107 x 175mm, / Weight : 0.370 Kg, / Gas supply Pressure : 280- 600 kpa with An integrated pressure resercur for the Stabilization of supplied pressure. Gasses option : O2 Air, End of scale Values :- 5L /Min, 10 L/Min, 15, 30, 40 L/Min	11
26	Pediatric BP Apparatus	With various cuff size	12
27	Pediatric Stethoscope	Pediatric size	06
28	Pediatric Resuscitation Kit	Handle, straight & curve Blade No.3,4,5 Ambu Bag 1 lit Oxygen reservoir	01
29	Neonatal Resuscitation Kit	Handle, Straight Carve Blade No.0,1,2 Ambu Bag 750 ml, Oxygen reservoir	01
30	Thermometer	Digital	60
31	IV Stands	Iron & Steel with Wheels	40
32	शिशोधारा यंत्र	2 lit Capacity Pot with Stand	06
33	Infantometer (Neonate)	100 Cm. Length	01
34	Vein Torch	Vein check	06
35	Tongue Depressor	Pediatric size Steel material	26

36	X-Ray View Box	<p>1. Ultra Thin X-ray Film illuminator using LED Lamps</p> <p>2. The Thickness should be 25mm – 30mm.</p> <p>3. It should be suitable for viewing 14" x 17" films.</p> <p>4. It should have LED lamps having life-span of more than 100,000 hours.</p> <p>5. It should have easy insertion & removal of the films.</p> <p>6. It should have homogeneous illumination & having luminance of more Than 1200 cd/m².</p> <p>7. It should have separate On-Off function & separate rotary continuous Adjustable brightness control for each panel.</p> <p>8. It should have auto on/off function with insertion and removal of films.</p> <p>9. It should have fully electronic continuous brightness control, with adjustment Range of approximately 90%.</p> <p>10. It should be directly connectable to power supply without any external adaptor</p> <p>11. It should have Flicker free high frequency light for reduction of eye strain.</p> <p>12. It should have external fuses for protection against power surge.</p> <p>13. Manufacturer/ Item should be ISO/FDA certified.</p>	20
37	E. T. Tube (Pediatric)	Size 2.5, 3, 3.5, 4	Each(10) 40
38	Induction Cooktop	- 2000 watt	40
39	Commode Chair	Heavy Duty Stainless Steel Foldable	12
40	Walker	Reciprocal folding walker	25
41	Adult wt machine	Standard Quality	27
42	Baby wt machine	Standard quality	04
43	Air Bed	Portable handly Dimensios-6x3x0.20 feet Man wt.-130 kg. (Capacity) Wt. Mattress 2-3 Kg. Inclusive of Air pump] Puncture kit / Repair Kit.	20
44	Double door Refrigerator	236 liter , inverter compressor, water dispenser, ice dispenser, double door type	09
45	Mixer grinder	500 watt mixer grinder with 3 jar	10
46	Wheel chair	Folding wheel chair and duel belt with breaks	25
47	Folding strature	Folding strature and duel belt , size 6 * 1.5 feet	15
48	Tripod for presentation	Maximum height extention 10 inches, max load capacity upto 2 lbs, rubber non sleep feet	01
49	Vaman Peet	Sitting Chair with basin attached with trolley 4*2=8	08
50	Kati Basti, Janu Basti, Hrid Basti yantra, Prishtavamsha	Plastic Material with different size.	16 set each 10

	Basti yantra		
51	Nasya Chair	Wooden square tube seat and back isd comfortable for the oatient to sit in semi recumbent position and head on the adjustable head support made of steel and woode.	06
52	Shirodhara Cap	Leather cap Small, Medium, and Large	14
53	Steel Coocker (15) litre atachment for steam (Nadi sweadan yantra)	Steel material, capacity 15 lit.	08
54	Dhum Pan yantra	Brass material	04
55	Basti Syringe	Steel material 200ml	12
56	Khal Stone (Hansraj Stone)	Ovel or Round shape of 15 inch diameter	02
57	Steam Bath	Deluxe Digital temp control, Timer height adjustable revolving stool, generator Manule water poring system	06
58	Massage Equipment	Fiber Glass Ancient Model tray, Legs with carving, Heavy with cushion 84"#30"	08
59	Sahlis Hb Tubes	Square Tubes of Standard company	20
60	New bar Chambers	Standard company	15
61	Rubber teat	Sooraj pharma	30
62	PH Papers pkts. multirange	Range 0 to 14	10
63	Multi stix	5 parameter	20 Box
64	Laboratory Autoclave (Fully Automatic)	Double walled construction fully made of SS. Top cover is die pressed to domed shape, supporting flange is made of thick SS flat entire joint Aragon arc welded & grounded to smooth but free finish. Supplied with two safety valves,pressure guage,steam release valve,& SS basket & drain valve at bottom. Digital temp.controller with inbuilt timer, low level cut offdevice, with auto purging system with solenoid valve.The autoclave is hydraulically tested for one & half times the operating pressure.Temp. range up to 121C/123C, operating pressure 15/1PSI 18"x24" (75lit)	01
65	Hot Air Oven	Laboratory type and memert type Double wall inside mild steel/aluminium /stainless steelouter chamber mild steel finishedwith powder coated ,temp. 250c /accuracy +-1c heating element of lab Laboratory oven of 500c to 2500c, 30x30x30cm,1250 watts	01
66	EDTA Vacutainers	k2 edta 2ml non vaccum blood collection tubespack 100pcs	100
67	Anlytical Physical Balance	Any standard company for Lab use	01
68	Syringe needle destroyer	Destroy needle of ty\pe 18 G, should have an power on and off switch, light weight , shock proof,	05

69	ELISA Reader & Washer	stainless steel 304 grade	01
70	ESR Reader	Any std company, fully automatic can store 35 test programs including graphs & washer –voltage 220+/- 10%volts Any standard company. Can handle 40 samples, analysis time 30min./60min. Working temp range-15-320C	01
71	Sample Rotator Machine	Any standard company. Silver coated RS-12 mini rotary shaker with 0-15 min timer fixed speed of 180 rpm	01
72	Electrolyte Analyser	Semi Automatic Type Bench top, Measure Na, K, Cl, Li	01
73	PH Meter	16*2 Alpha , Numeric Display, Lcd Type , Range 0 To 14 Upto 100 sampals	01
74	Laminar Flow Chamber(Bio safety Laminate)	max working temprture upto 1160 digree C Power Supply 230 Volts, 9*4*4 outer 18*14*4	01
75	laringo scope with Bled paid/adults	Stainless Steel Quality	1 each
76	Patient carrying Trolley	Stainless Steel Quality	06
77	Sterilizer vertical double dekar	Electricals High pressure standard quality	06
78	Urethral Dialitars	Stainless Steel isorted sizes Quality	2 set
79	Roti maker	staniless steel (200) Roti capacity	01
80	Sahli's Hb tube	Square tube;	50
81	Hb pipette	Zero eirr 20cmm pipette	20
82	WBC Pipette	Imported/ Indian	20
83	Dropper	Made of latex with rubber teat	20
84	Red cell pipette	Imported/ Indian	20
85	Wintrobe's tube	Made of glass	20
86	Pasteur's pipette	Made of glass	50
87	Westergrens pipette	Made of glass	50
88	Westergrens's stand	For three tubes	10
89	Urinometer	Routine upto specific gravity 1.02	40
90	Ultraviolet lamp	Regular type	02
91	Thermometer	Digital	50
92	Physical balance	250 gms, sensitivity 2-3mgm with 3 stone knife edge agate fitted in teak wood glass case	01
93	Bunsen burner	With stop cock to control on/off	50

94	Sterile vessels/bottle to collect samples	with EDTA	100
95	Torch	With LED light, good focus	02
96	Measuring Tape	Latex material	02
97	Reflectors(Mirrors)	With LED light	01
98	Nasal speculum	Made of stainless steel set of three	02
99	Laryngoscope	With LED	02
100	Catheters	Made of rubber	04
101	HBs Ag kit	With standard control	01
102	HIV kit - Tridot (method by T Mitra)	With standard control	01
103	CT and BT kit	With capillaries and filter paper	10
104	Renal profile, LFT kit, Lipid profile, Blood Sugar kit	With standard control	01
105	Sterile disposable lancet/needle	24 Gauze	10pks
106	Glass rod	Stirring rods 100mm,3mm pack size 100n 200mm,3mm pack size 100n 200mm,4mm pack size 100n 200mm,6mm pack size 100n	01pk 01pk 01pk 01pk
107	Cover slip	2mm square(pkt of 10 gms)	10pks
108	Cleaned slides	Size 76 mm*26mm pkt of 50 slides	10pks
109	Litmus paper	Red and Blue	10pks
110	Test tube	12*75 mm Borosilicate glass 12*100 mm Borosilicate glass 15*125 mm Borosilicate glass	100 100 100
111	Separating funnels of various sizes	125 ml Borosilicate glass 250 ml Borosilicate glass 500 ml Borosilicate glass	10 10 10
112	Capillary Tubes	Melting point microcapillary	10

113	Magnifying lens	75mm, one endsealed pkt of 100 capillaries	box
114	Multi stix	With handle 4" dia.	10
115	सिलींग मशीन	5 parameter	10 box
		१८ इंच बेग पैक करण्याची क्षमता	1