

अधिष्ठाता यांचे कार्यालय,
शासकीय आयुर्वेद रुग्णालय, वजिराबाद, नांदेड.

OFFICE OF THE DEAN,
GOVT. AYURVED HOSPITAL, VAZIRABAD, NANDED.

Email ID :- gahnanded@gmail.com दूरध्वनी क्रमांक :- (०२४६२) २३२८६७ फॅक्स क्र. २३४०२६

जो. क्र. शा.आ.रु.ता./श्रीपद्मीभांडार/

/२०२१

दिनांक :- / /२०२१

प्रति,

विषय :- शासकीय आयुर्वेद रुग्णालय, नांदेड.

यंत्रसामुग्री/उपकरणे इत्यादी पुरवठा करण्याकरिता दरपत्रक सादर करणे बाबत.

उपरोक्त विषयाच्या अनुषंगाने आपणांस कळविण्यात येते की, जिल्हा वार्षिक योजना (सर्वसाधारण) २०२०-२०२१ अंतर्गत मंजूर निधी शासकीय आयुर्वेद महाविद्यालय व रुग्णालय, नांदेड यांचे वळकटीकरण तथा श्रेणीवर्धन याकरिता मंजूर निधीतून यंत्रसामुग्री/उपकरणे खरेदी करण्याकरिता दरपत्रके मागविण्यात येत आहेत. सदर यंत्रसामुग्रीचे नावं, नग व विनिर्दिष्टे (Specification) सोबत जोडण्यात आले आहेत.

उपणांस सुचित करण्यांत येते की, रुग्णालय व महाविद्यालय करिता आवश्यक यंत्रसामुग्री/उपकरणे इत्यादी पुरवठा करण्याकरिता सोबत जोडण्यात आलेल्या विनिर्दिष्टेनुसार (Specification) खालील नमुद यंत्रसामुग्रीचा पुरवठा करण्याकरिता कमीत कमी दराची मोहोरबंद पाकीटातील दरपत्रके दिनांक ०८.०२.२०२१ रोजी सायं ०५:०० वाजेपर्यंत पोहोचतील अशा बेताने अधिष्ठाता यांचे कार्यालयात खालील नमुद अटी व शर्तीच्या अधिन राहून सादर करण्यांत यावीत.

अ.क्र.	बाब	कालावधी
१	दरपत्रक सादर करण्याचा कालावधी	दि. २९/०१/२०२१ ते ०८/०२/२०२१ (या कार्यालयीन कामाकाजाच्या दिवशी व वेळेत शासकीय व स्थानिक सुटीचे दिवस वगळून सकाळी १०-०० ते सायं ०५-०० वाजे पर्यंत)
२	दरपत्रक उघडण्याचा दिनांक व वेळ	दि. ०९/०२/२०२१ रोजी सकाळी ११-०० वाजता
३	दरपत्रक उघडण्याचे ठिकाण	अधिष्ठाता शासकीय आयुर्वेद रुग्णालय, नांदेड यांचे दालन.

अटी व शर्ती

- दिनांक ०८/०२/२०२१ रोजी सायंकाळी ०५:०० वाजल्यानंतर प्राप्त होणाऱ्या दरपत्रकाचा विचार केला जाणार नाही.
- दरपत्रक सादर करतांना अशा प्रकारच्या खरेदीवर शासनाकडून आकारल्या जाणारे विविध कर (टॅक्सेस) स्वतंत्रपणे नमुद करावे.
- दरपत्रक सादर करतांना जी.एस.टी. क्रमांक नमुद असलेल्या लेटरहेड वर सादर करावे.
- ज्या यंत्र सामुग्री करिता दरपत्रक सादर करण्यात येत आहे. ते आय.एस.आय. चिन्हांकित असेल तर ठळकपणे नमुद करावे.
- दरपत्रक सादर करणाऱ्या दरपत्रकाचा किंवा संस्थेच्या दरपत्रक प्रक्रियेत भाग घेणाऱ्या अन्य संस्थेमध्ये थेट स्वारस्य असु नये तसे आढळल्यास दोन्ही किंवा अधिक दरपत्रके नाकारण्यात येतील. याकरिता सोबत जोडण्या आलेली हमी पत्र दरपत्रक धारकास सादर करणे बंधनकारक राहिल. (सोबत हमीपत्र विहित नमुण्यात सादर करावेत)

६. यंत्रसामुग्री खरेदी पश्चात यंत्रसामुग्री करिता देण्यात येणारा हमी कालावधी (वारंटी) काळापणे नमुद करावा.
७. दरपत्रका सोबत ज्या यंत्रसामुग्री करिता दरपत्रक सादर करण्यात येत आहे, त्या यंत्रसामुग्रीची विस्तृत माहिती असलेली माहिती पत्रक (Products catalogue) जोडावे.
८. दरपत्रक सादर करतेवेळी मोहोरबंद/सिलबंद पॉकीटावर यंत्रसामुग्री करिताचे दरपत्रक असे स्पष्ट करावे.
९. खरेदी प्रक्रिया यशस्वी दरपत्रक स्विकारण्यापुर्वी किंवा यशस्वी दरपत्रक स्विकारण्यात आल्यानंतर देखील कोणत्याही वेळी लेखी नमुद करावयाचा कारणामुळे रद्द करण्याचे अधिकार हे कार्यालय राखुन ठेवीत आहेत.
१०. विहित कालावधीत प्राप्त दरपत्रके दिनांक ०९/०२/२०२१ रोजी सकाळी ११-०० वाजता अधिष्ठाता यांच्या दालनाल उघडण्यात येतील त्यावेळी कंपनीच्या एका प्रतिनिधीस स्वखर्चान उपस्थित राहण्यास मुभा राहिल.
११. दरपत्रक उघडल्यानंतर आपण सादर केलेल्या यंत्राच्या विनिर्दिष्टे (Specification) नुसार सदर यंत्राची प्रत्यक्ष पडताळणी व आवश्यकता वाटल्यास चाचणी रुग्णालयातील संबंधित तज्ज्ञामार्फत करण्यात येईल त्याकरिता आपण आवश्यक ती व्यवस्था करुन ठेवावी.
१२. पुरवठादाराने सदर खरेदी संदर्भातील आवश्यक ती माहिती या सोबतच्या Annexure-B मध्ये सादर करावी.
१३. प्रत्येक यंत्रसामुग्री/उपकरणे करिता स्वतंत्र दर पत्रके सादर करावीत.
१४. खरेदी प्रक्रियेचा अटी व शर्तीमध्ये कोणताही बदल करण्याचा अधिकार अधिष्ठाता शासकीय आयुर्वेद रुग्णालय, नांदेड हे राखुन ठेवीत आहेत.



(डॉ.वाय.आर.पाटील)
अधिष्ठाता
शासकीय आयुर्वेद रुग्णालय, नांदेड.

सहपत्र (Annrxure)- A
शासकीय आयुर्वेद महाविद्यालय, नांदेड.
 SPECIFICATION OF.....

अ. क्र.	दरपत्रकाब्यारे आवश्यक असलेले विनिर्दिष्टे Quotation as asked in the quotation	पुरवठादाराने प्रस्तावित केलेली तत्सम विनिर्दिष्टे कंपनीच्या नावासहित Equivalent Specifications quoted by the supplier with the name of the manufacturer	पुरवठादाराने नमूद केलेली विनिर्दिष्टे दरपत्रकान्वये विचारणा करण्यांत आलेल्या विनिर्दिष्टानुसारच आहेत किंवा नाही. प्रत्येक बाबीसमोर होय किंवा नाही स्पष्टपणे नमूद करावे. Whether the Specification quoted by the supplier is as per specification asked for indicated against each item clearly YES/NO	पुरवठादाराने नमूद केलेली विनिर्दिष्टे दरपत्रकान्वये विचारणा करण्यांत आलेल्या विनिर्दिष्टे यामध्ये काही बदल वा तफावत असल्यास प्रत्येक बाबीसमोर स्पष्टपणे नमूद करावे. If the Specification quoted is not a per specification then variation /deviation should be clearly indicated against each item asked for.
१.				

Annexure-B

कंपनीच्या लेटरहेडवर खालील नमूद तपशिलाप्रमाणे माहिती भरणे.

प्रति,
अधिष्ठाता
शासकीय आयुर्वेद रुग्णालय,
वजिराबाद, नांदेड.

जिल्हा वार्षिक योजना (सर्वसाधारण) - २०२०-२०२१

या यंत्रसामुग्री करिताचे दरपत्रक

अ. क्र.	तपशिल	आवश्यक माहिती वा प्रमाणपत्रे जोडण्यात आलेली आहेत किंवा कसे
१	पुरवठादाराचे नाव	
२	कंपनीचा पत्ता, दुरध्वनी ई-मेल आयडीसहीत	
३	कंपनीचा नोंदणी क्रमांक (Registration No.)	
४	यंत्राचे नाव आणि मॉडेल क्रमांक	
५	विनिर्दिष्टे (सोबत जोडण्यात आलेल्या सहपत्रानुसार) Annexure-A	
६	G.S.T. No.	
७	शासनाचे कर	
८	इतर कर	
९	I.S.I Mark इ(आयएसआय चिन्हांकित आहे किंवा कसे ? असल्यास तत्संबंधीचे प्रधिकरणाकडील प्रमाणपत्राची छायांकित प्रत जोडण्यात आली आहे. किंवा कसे ?)	
१०	उत्पादक असल्यास तसे नमूद करावे अथवा प्रधिकृत पुरवठादार असल्याबाबतचे उत्पादक कंपनीचे पत्र.	
११	सदर यंत्राचा इतर शासकीय/निमशासकीय कार्यालयास पुरवठा केला असल्यास अशा कार्यालयाचे पत्ते व दुरध्वनी क्र. अथवा पुरवठा आदेशांची छायांकित प्रत.	
१२	हमी कालावधी (Warranty Period).	
१३	यंत्राची माहिती पुस्तिका/पत्रक	
१४	हमीपत्र	
१५	इतर अटी व शर्ती असल्यास	

स्थळ :-

ठिकाण :-

पुरवठादाराचा अधिकृत स्वाक्षरी व शिक्कांसहित

हमीपत्र - १ (Undertaking)

मी M/s

पत्ता :-

लेखी हमी देत आहे की, आपल्या कार्यालयामार्फत

..... खरेदी करणेसाठी राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी भाग घेत असून माझे किंवा माझ्या संस्थेचे खरेदी प्राधिकारी / खरेदी समिती वा कार्यालयाशी कोणत्याही प्रकारचे हितसंबंध / संघर्ष नाहीत तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ :-

पुरवठादारांच्या अधिकृत स्वाक्षरी व शिक्कासह

दिनांक :-

हमीपत्र - २ (Undertaking)

मी M/s

पत्ता :- लेखी हमी देत आहे

की, आपल्या कार्यालयामार्फत खरेदी करणेसाठी

राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी/माझी संस्था केवळ एकच दरपत्रक सादर करित आहे.

त्याचप्रमाणे दरपत्रक प्रक्रियेत भाग घेणाऱ्या अन्य संस्थासोबत माझा/माझ्या संस्थेचा कोणत्याही प्रकारचा प्रत्यक्ष

/अप्रत्यक्ष संबंध वा सहभाग नाही तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात

यावे.

स्थळ :-

पुरवठादारांच्या अधिकृत स्वाक्षरी व शिक्कासह

दिनांक :-

**GOVT. AYURVED HOSPITAL, VAZIRABAD, NANDED.
PURCHASING OF INSTRUMENTS & MACHINERY UNDER (D.P.D.C) BUDGET 2020-21**

Sr.No	Name Instruments & Machinery	Specification	Qty,	Rate Per Unit
1	Cautery Machine	Microprocessor based IGBT Technology Tissue Response Technology CE, ISO 2001:2008, ISO 13485:2003, IEC 60601-1 High Crest factor Cardiac protection circuit Auto on facility for coagulation Endo cut facility Separate Monopolar / Bipolar system (Two different footswitch for both monopolar & bipolar for simultaneous use of machine by two user) Louvers & fans for heat dissipation Any accessories compatible (Reusable accessories more prefer) Upgradation facility Less Thermal spread Component level machine manufacturing to avoid problem occurs in SMD level technologies due to service point of view Onsite repairable , component level servicing Stand by machine network at local area Lowest servicing price Local service center at local area since 5 years 1 Watt step for calibration Particulers max power (watt) pure cute 400 w Bland 1 250 w Blend 2 200 w spray-120 w force-150 w	1	

		<p>soft-100w Biopolar force coag -80 w Biopolar cut-80 w Operating frequency 300 Khzmint Weight-6.8 kg Rimoundion-300><140><450+</p>		
2	Vertical Autoclave Single drum	<p>Internal capacity -22 Litre Load- 2.00kw chamber dimension -10"x18" working pressure 1.2 kg f/cm 59. g(15 psi-17 psi) sterilizing 121°c to 134. c (150. c-optional) heat average-<=±1°c sterilization Time-25 to 30 Minutes digital timer - 0 to 99 min construction -double/triple wall & ss4mp construction outer chamber (304 stainless steel) inner chamber -304ss insulation-glass wool waste container - stainless steel Door -Lid - foot+lifted with safety & interlock device door locking type -radial type safety features radial locking, safety valve, low water detector & pressure interlock power supply -220 volts/440 volts.</p>	1	
3	Stainless steel Dressing Trolley	<p>length-1080 mm width 490mm height-840 mm frame - 25 mm x25mm ss square tube frame with two stainless steel shelves with railing on top shelf Trolley mounted on 100mm dia castors ss bowl 360mm diameter ss bucket 300 diameter</p>	1	
4	Examination Table	<p>Size -72"x24"x30 "ht Top design in two section one top panel fixed with coach & another head panel for head low & head up position on each assembly tops are made in M.S. surgery gauze sheet coach providers three drawers with lock & three lockers provide locking system with heavy handle attached one single step providing at below total coach expiry powder coated in colour combination.</p>	1	
5	Infant/neonatal Warmer with ressuscitation unit	<p>1. German made imported ceramic heater with p parabolic reflector microprocessor base control temperature controller , air and manual control with LCD Screen , high temp sensor for high temp cut off at 39 degree .Swivelling source box for x-ray .Oxygenation - inbuilt flow suction apparatus with pressure control. Incubation providing complete</p>	1	

		set of resuscitation items .Baby tray four sides clear acrylic panels with head up and down facility Lower three drawer facility		
6	Electrical Cervical & Lumbar Traction Machine (LCD)	Main supply 220V r0 HZ .Consumption - 85 watts Lumbar Force 15-45 kg with doublers upto 90 kg .Cervical force - 3,4,5,6,7,8,9,10,11,12,13 kg.Model selection -Separate switch . Timer - Digital (0-99 min) with AUTO Alarm.Hold time - 20,40,60,80 sec.Rest TIME 1,5,10,15,SEC .Weight 10 kg .Dimension (H#W #L) - 30#31#28 CM	1	
7	B.M.I. Machine	B.M. I. Machine 24 col.dot Matrix with thermal printer can print date, time, weight ,height, B.M.I. Machine should have - Voice input 1. Self weight - approx 30 Kg 2. Self height - 250 cm (approx) 3. Operating temp -0° to 50° 4. Power supply - 250 V.A.C. Should be supplied with adequate cables wires 5. weighing capacity - 200 Kg 6. weighing accuracy - upto 100 gm 7. Height range -(min) 50 to 200 cm 8. Height accuracy -(min)0. 5cm 9. Plat form size- (max) 360 x 300 mm 10. Guarranty -(min) 1 year 11. Annual Maintenance- upto next four years	1	
8	Spirometer with printer	Examine lung function related parameters for patients. Accuracy 85-100% Alarm system: Low & High Model: SP-100 , Printed report, PC based spirometer along with latest window version with adequate aauthentic software	1	

9	Massage table with attachments	Stainless steel 304 grade, pine wood and polycarbonated Height :-30" Weight :- 85 kg. Measurement :- Length 7 feet width 3 feet height 2.5 feet. Weight bearing capacity 150 Kg Attachment :- 1 Attached vaman peeth.2. Attached shirodhara stand 3. Attached sarvang dhara stand 4. Attached basti stand. 5. foot rest 6. traction. 7. Uttarbast.	1	
10	Glucometer		1	
11	OT care or Fumigation machine	Fumigation machine with 304 stainless steel body, Aerosol disinfectant fumigator, pre fitted liquid flow control valve, portable, 5 lit tank capacity, 230 V, 50 Hz,	1	
	NIBP multipara Monitor	5 Parameter – SPO2, ECG, NIBP, RESPIRATION, TEMP, LCD display, ECG lead mode – 5 lead, mouse, keyboard operation applicable, ultra quiet fan less design, multiple working modes option, NIBP assisted vein puncture. Display – 12.1 inch, HR range 15-380 bpm, suitable for neonates and paediatrics, alarm indicator for ECG, SPO2, NIBP, RESPIRATION, RESP 2-temp and Arrhythmia indicator, charge lite battery 500. NIBP measurement data can be stored and recall, reusable, NIBP Cuff pediatric size 3 sets	1	
12				
13	Syring pump	Automatic detection of syringe size capacity, Flow rate can be change during infusion, 3 level, Occlusion alarm limit setting, Bolus facility for fast delivery of drug, Automatic rate calculation with volume and time limit, Syringe 5, 10, 20, 30, 50, 60cc, 9.6 volt, 1300mAm, Battery backup 5-6hrs/ 2.3kg, 306x135x127mm dimensions.	1	
14	Warmer with LED double surface phototherapy	Photography latest high bright super flex LED double surface. Phototherapy with digital timer surface photography with digital timer german meck ceramic heater with parabolic reflector unit made with high quality FRR fiber for safety clear acrylic base for bottom surface phototherapy with Xray cassette compartment, lowerty for utility, mounted anti skid caters with locking facilities, owl space baby tray with three side collapsible acrylic walls both side monitor (bracketed, 1 V stans to mount multiple syringe pumps.	1	

15	Nadi swedanayantra	Five litre stainless steel heavy bottom cooker with 4 meter rubber tube.	1	
16	Stadiometer	Measure height from 0 to 180cm.	1	
17	Manual BP apparatus	Mercury	1	
18	Digital BP apparatus	Fully automatic, operating oscillometric principle, pediatric cuff with IntelliSense technology, body movement detection, hypertension indicator, irregular heart beat detection.	1	
19	Panchakarma table with shirodhara yantra	Standard wooden specialised table with steel stand with pot attachment, Metallic shirodhara pot with capacity 1-2 litre with chain attachment. Wooden table with curved edges of size 6 ft x 2.5 ft with standard height, four corner side holes for drainage of medicated oil.	1	
20	Pediatric Resuscitation Kit	Steel body laryngoscope with straight and curved blades no 2,3 and 4, AMBU bag 750ml capacity with appropriate masks, ET tubes no.4, 4,5,5 and 6 one each.	1	
21	Otoscope (paediatric)	Fiberoptic otoscope compact pocket, colour blue, replaceable and rechargeable batteries with mini Nt table charger, 4 reusable tips and 10 disposable tips, size 12X7X4 cm	1	
22	Autoclave drum	Stainless steel, Size 11"x9" – 1 unit, 8" x 9" – 1 unit.	1	
23	Bed pan	Stainless steel, Non Rusting, Non magnetic	1	
24	Curtain Stand	3 face exam room privacy screen, Stainless steel body, with wheels at bottom, folding facility of height 60 inch*24inch size	1	
25	Electric Weiging machine	Electronic digital weighing Platform scale with stand, weighing capacity upto 200 Kg	1	
26	Instrument cabinet	Non Rusting, Non magnetic, size = 960x400x1750mm epoxy coated mild steel frame work stainless steel bottom.	1	

27	Anesthesia Trolley	Stainless steel -casters wheel with lock -650x120x1020 mm -8 drawer -1 sliding shelf -1 waste basin -1doughe can -oxygen holder stand	1	
28	Enema canne	Plastic body with capacity of 1 liter, with 1m plastic tube and nozzles of various shapes	1	
29	Obstretic Table	72*22*30 LWH diamentions, 3 section cushioned top with stainless steel bowl, Height adjustable, pair of knee crutches, back rest, seat adjustable by hand lever	1	
30	Sterilizer	30 cm x15cmx12.5cm -Electric supply of200 watt -with hinged lid ,fitted with immersion rod & water outlet tap -Stainless steel body -Thermostatic & electric shock preventing system -Tray lifting system	1	
31	Foetal Doppler	Portable faetal doppler with clearly sound 2 Mhz. probe with LED disply, Faetal Heart rate, battery backup system/ artificial battery input system in case of electricity failure	1	
32	Electric Steam Inhaler	Plastic body, of any standerd quality	1	
33	Cardiac Monitor	39*27*29 cm in HWB, 4 KG Weigh. advanced steem line outline design inbuilt handle ,monitor with adult paediatric use disply of SpO2, ECG, NIBP, Pulse.	1	
34	Kadhai (Iron) 22cm x 22cm	Iron material, of standerd of size (22cm x 22cm)	1	
35	Kadha Galani (Steel) 6 No.	Stainless steel Galani of standerd quality	1	
36	Mixer Heavy Duty	550 watt, 230 volt , 3 speed control, multipurpose, stainless steel lead, with 3 pot set	1	
37	Inverter 180 (12 V) with battery UPS	180 VA (12V) Invertor with Battery	1	

38	Bipolar Ophthalmic Cautery	Radiofrequency- 455-515KHz, output RF power-5-7W, input supply-220V/50Hz, reusable forceps, erasers, cables, foot switch control, audible indications, 0-9 level variable power output, test certificate and instruction leaflet	1	
39	Schioltz tonometer	All vital parts are made up of aluminium other parts are BRASS chrome plated complete in case. Perfect reading of scale with rd pointer supplied with 3 weight (5.5gm, 7.5gm, 10gm), with conversion table. Precision measurement on a scale of 0-20subdivisions, and 0 to -1 subdivisions, high quality agate bearing	1	
40	Chalazion clamp	Stainless steel medium-2 Large 2	1	
41	Chalazion scoop/ currete	Stainless steel, small-2 Medium-2	1	
42	Foreign body spud	Stainless steel, pointed tip	1	
43	U.V light instrument sterilization cabinet	A high output , 10,000 hr rated, 8 watt, 254mm, ultraviolet bulb of size (G8T5,2pin, 12" long), grounded power cord (220V, 7' long)	1	
44	Draffin's bipod stand with plates	stainless steel	1	
45	B.P. Handle	stainless steel No-7 - 2 No.4-2	1	
46	Binocular Microscopes with Oil Immersion	Body Interchangeable, inclined binocular body, 360deg rotatable head, Eyepieces-Highest quality 10X wide angle antifungal eyepiece, objectives parfocal, antifungal coated 4X, 10X, 40X & 100X(oil Immerssion) with plan achromatic correction. Optical system-infinity corrected, Stage Horizontal mechanical stage, preferably 100 x 140 mm with fine vernier graduations designed with convenient coaxial adjustments for slide manipulation preferably through 30 x 70 mm. Sub stage-Abbe condensor focusable continuously variable iris diaphragm.	1	
47	Haematocrit Tube	Blood Sedimentation Tubes of Borocillate glass Suitable for determining sedimentation rates & Volume of Packed cells, the values stated in SI units. ragarded as standards.	1	

48	Stop watches	Should be of ideal watch for meaning short or medium intervals it consist of stainless steel single metallic, balances nickle based allow hairspring, anchor escapmet & with 13 jewels fixed in the structure over all size should be 70x50x16mm weight -120 gms.	1	
49	Tuning Forks	Preferable any standard make for laboratory use.	1	
50	Pressure Cooker (5) Ltr.	ISI mark	1	
51	Pressure Cooker (15) Ltr.	ISI mark	1	
52	Aluminium Pot	Standard quality capacity in kg (2,5, 10, 12)	1	
53	Mixer grinder Commerical	isi mark standard quality	1	
54	Barracating strip stand	Stainless steel	1	
55	Jumbo oxygen carrying trolley	Stainless steel	1	
56	Aluminium Ladder (15 Fit)	Standard quality	1	
57	Goods transport trolley	Stainless steel	1	
58	Kitchen Trolley	Stainless steel	1	
59	Digital sterilizer (20 Inch)	Standard quality	1	
60	Floor cleaning machine for office	Operation mannual weight -45 kg brush & paddiameter -430mm.	1	
61	Sanitizer sensor	Automatic Contactless Hand Sanitizer Dispensor (5/10 Ltr.)	1	
62	Revoling stool	Stainless steel	1	
63	Exhaust fan	Standard quality isi mark	1	
64	Emergancy Tournch	Standard quality with Charging	1	
65	Room tent Cooler (Industrial)	Fibre body 60-70 Ltr water capacity	1	
66	Steel Glass	Stainless steel	1	
67	Wall Watch	Standard quality	1	
68	Bowl Stand	Stainless steel	1	
69	Kidney tray	Stainless steel	1	
70	Moveble TV Cabinet	53 inch size Cabinet	1	
71	Steel cupboard	heavy Gauge 6*3 size	1	
72	Sound System with amplifier	Standard compny	1	
73	steel Pan with cover Plate	size 1 kg, 10 nos .2kg, 10 nos .2.5kg, 10 nos .5kg, 6 nos	1	
74	Alluminium Pan with cover plate	size 1 kg, 10 nos .2kg, 10 nos .2.5kg, 10 nos .5kg, 6 nos	1	
75	steel Ladle	Stainless steel	1	
76	Regulor spoone	Stainless steel	1	

77	Steel Plate	Stainless steel		1	
78	Knife	Standard quality		1	
79	steel Bowl	Stainless steel		1	
80	Sieve the flour with plates	Standard quality		1	
81	Tea strainer	Standard quality		1	
82	granaries	Stainless steel 100 kg ,50 kg ,20 kg		1	
83	long frying pan	iron long frying pan		1	
84	steel basket	Stainless steel with cover plate		1	
85	Steel reversals	Standard quality		1	
86	Beehive (dike)	Stainless steel, Wooden		1	
87	Cheat Rolling pin	Wooden rolling pin		1	
88	Steel fork tweezers	Standard quality		1	
89	Plastic Bins	Standard quality 1kg size		1	
90	Steel Bins	Stainless steel		1	
91	Steel Bucket	5 ltr, 10 ltr with cover plate		1	
92	Big size plate (Aluminium)			1	
93	Surviving Bowl	Stainless steel		1	
94	Steel Egg Tray	Stainless steel		1	
95	Oil Kettle	Stainless steel ,size 3 kg		1	
96	Tea Thermos	Stainless steel 3 ltr		1	
97	Milk Measurement Cup	100 gm,200gm, 1 ltr		1	
98	Electronic Weight Machine with Display	ISI mark 100 kg,		1	
99	Butter Milk Ravi	Stainless steel		1	
100	LED Board Smart Board	Interactive LED Smart board (LED flat panel) Display area - 1895x1066 mm 3840x2160 Life time :- >30000 hours next generation anti glare glass Infrared touch.	size-86" Resolution :- Display colours - 10 bit Touch surface Touch sensor -	1	
101	American Thali	Stainless steel, molded 5 compartment thali.		1	

102	Microscope	clinical trinocular microscope with battery backup body Aluminium die-cast body. Focusing system- co-axial coarse & Fine fine control s coarse focus range-20 mm Condensor - Abbe focous range - 0.25mm condenssor with perture iris diagram illumination- 3 w LED 4x, 0.10 / 18.5m 10X 0.25/10.6mm loaded) 0.65/0.6mm loaded)1.25/0.13mm accessories- Camera 40X(spring 100X (Oil sping Accessories- Camera Objectives,	1	
103	Digital Camaera for Mocrscope	Digital camera for magnus trinocular Led microscope model- NX21i-TR with 0.5XTR adapter.	1	
104	Chapati Bhatti with two Burner Stove in S.S. & Exhaust Chemni	Heavy Gauge Steenless Steel (Modular)	1	
105	Modular Kitchen Storge & Modular washbBasin in S.S.	Heavy specify Gauge Steenless Steel (Modular)	1	
106	Biomtric Machin (Iris/Face)	Standers Quality (Iris/Face Recognition Detection)	1	
107	Foot Streps	Two Stapes size, Steel, Painted, 6" 2x2" wide	1	
108	Rotary microtome	Section thickness setting -05-60 μ mn from 0.5 μ mn -2 μ mn in 0.5 μ mn increment from 20ml -20 μ mn in 1 μ increments from 20 μ m-60 μ in 5 μ mn in i5 μ mn increments Vertical strake - 60 mm specification orientation - horizontal 8', vertical 8' Object feed - 28 mm Max sectiony area - 602 mm x 50 mm microtone should be supplied complete as above with disposable blade holder for both low & high profile blades, a sample pack of 10 each disposable blades & universal specimen clamp & universal cassette clamp & 10 sample cassettes.	1	

(Dr. Y.R.Patil)

Dean

Govt. Ayurved Hospital, Nanded.